


At about 30 minutes, remove the top lid completely to take a picture using the KinTouch App.




Confirm and save the test results according to the instruction in the App.




[RESULT INTERPRETATION]

Red in both windows.



POSITIVE 1st or 2nd Degree Relatives

Red in the right window. Yellow in the left window.



NEGATIVE Neither 1st nor 2nd Degree Relatives.

Yellow in the right window.



INVALID
Uncertain Results Due to
Control Failure.





Cover the cassette firmly with the top lid.

७10min⋅⋅⋅

After ten minutes, read the color pattern of the windows.

interpret the test results according to the following rule.





Break open a Water Packet along the cutline in the spout

Open the cassette and place the spout firmly against the middle (M) touchpad. Let water flow onto the touchpad naturally with a gentle squeeze.

It takes about one minute to empty the entire Water Packet and the bands in the two windows will start to change colors.



DO NOT OVER-SQUEEZE





20sec



Open the cassette and press the middle (M) and right (R) touchpads gently with moist index and middle fingers for 20 seconds.

Place the top lid back to the cassette to prevent contamination.



05





SUBJECT #2:

Wash hands thoroughly and dry with a fresh paper lowel.

Grab a fresh Cleansing Wipe with the left-hand thumb, index and middle fingers and clean these fingers briefly. Discard the used wipe.

Rub the thumb against index and middle fingers for 5 seconds.

Keep fingers moist for the next step.



5 SECONDS





40sec

De Jamil

Open the cassette and press the left (L) touchpad gently with the moist index finger for 40 seconds.

Place the top lid back to the cassette to prevent contamination.



03





SUBJECT #1:

Wash hands thoroughly and dry with a fresh paper towel.

Grab a fresh Cleansing Wipe with the left-hand thumb and index fingers and clean these fingers briefly. Discard the used wipe.

Rub the thumb against index linger for 5 seconds.

Keep fingers moist for the next step.



5 SECONDS



Read the How-to Guide thoroughly before taking the test.

- Remove the Cleansing Wipe (2 packets) from the seat
- Move a cassette to the same spot.
- Unlock the top lid of the cassette and keep its lid loosely on.

